Official copies of these procedures are maintained at this website.

Before using a printed copy, verify that it is the most current version by checking the document issue date on this website. Signed copies of these official procedures are maintained at the Training Office.

C-A OPERATIONS PROCEDURES MANUAL

ATTACHMENT

8.15.1.a Chipmunk Calibration Log Sheet

| | C-A OPM Procedu | ires in wh | nich this Attachm | nent is used. | | |
|---------|-----------------|--|-------------------|-----------------|--|--|
| | 8.15.1 | | | | | |
| | | | | | | |
| | <u>Ha</u> | nd Proces | ssed Changes | | | |
| HPC No. | <u>Date</u> | | Page Nos. | <u>Initials</u> | | |
| | | _ | | | | |
| | | _ | | | | |
| | | _ | | | | |
| | | | | | | |
| | | | | | | |
| | Approved: | Signature On File Collider-Accelerator Department Chairman | | | | |

V. Castillo

CHIPMUNK CALIBRATION LOG SHEET

| chnician Name | | | Date: | | |
|--|------------|------------------------------|------------------------------------|--|--|
| strument S/N: | | | | | |
| te of last calibration | : | T | Test Box calibration date: | | |
| 30 m Ci So | ource Dis | tance @ 8 mr/hr: | cm, Source # 1595 | | |
| | | tance @ 80 mr/hr: | | | |
| 150 m Ci Sc | ource Dist | tance @ 150 mr/hr: | cm, Source # S-656 | | |
| uality Factor: _ | | Failsafe: | minutes (1.5 to 4 min. acceptable) | | |
| wered ≥60 mins.(1hr | | | • | | |
| | | As Found | As Left | | |
| 1000 sec background count accept. range 50-200 cts | | counts | counts ^ω | | |
| 100 sec 8 mr/hr accept. range 432-477 c | | 1 counts | counts | | |
| | | 2. counts | counts | | |
| | | 3 counts | counts | | |
| Average | | counts | counts | | |
| 100 sec 80 mr/hr | 1 | | 100 sec 150 mr/hr counts | | |
| accept. range 4090-5000 | 2 | counts | t. range 7670-9375 counts | | |
| | 3 | counts | counts | | |
| Average | | counts | count | | |
| S Level: | | | | | |
| D | | : 1 1:-4 - £41 | - J). | | |
| Remarks (include | апу гера | ir and list of parts replace | ea): | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

| Signature: | | |
|------------|---|--|
| | (Signature on person(s) performing the calibration) | |

 $^{\mbox{\tiny ω}}$ If "as found" background is in acceptable range, "as left" background need not be repeated.